



Wigam S.p.A.
Loc. Spedale 10/b
52018 Castel San Niccolò
(AR) Italia
Tel. 0575 / 5011
Fax. 0575 / 501200



Declaration of Conformity

We, signers of this declaration, declare under our own exclusive responsibility, that the digital vacuum gauge, model:

VG 64-LX

Are designed according to the following directives prescriptions:

- 2014/35/UE Directive on low voltage
- 2014/30/UE Directive on electromagnetic compatibility,

And according to the following directives EMC:

- EN 55011/EN55022 Level B
- EN 61000-4-2 : 2009 Level A
- EN 61000-4-3 : 2009 Level A

Castel San Niccolò
05/12/2018

Gastone Vangelisti
(President)

A handwritten signature in black ink, appearing to read 'Gastone Vangelisti'.

TK109 Fiber glass insulated thermocouple

Code	Model	Measuring range	Cable length	Dimensions	Weight
09012013	TK109	-50 to +200°C	3 m	100 x 80 x 10 mm	0,02 kg


Calibration

Code	Description
09053005	Calibration report ref. Accredia for TK109


TK109-STRIP Fiber glass insulated thermocouple with strip

Code	Model	Measuring range	Cable length	Dimensions	Weight
09012037	TK109-STRIP	-50 to +200°C	3 m	100 x 80 x 10 mm	0,02 kg


Calibration

Code	Description
09053005	Calibration report ref. Accredia for TK109-STRIP

Strip suitable for hoses max. diameter 80 mm


TK104 Fast reacting probe for air and gases

Code	Model	Measuring range	Cable length	Dimensions	Weight
09012019	TK104	-50 to +800°C	27 cm	380 x 100 x 10 mm	0,12 kg


Calibration

Code	Description
09053001	Calibration report ref. Accredia for TK104


TK106 Insertion probe for food-stuff, plastic, rubber etc.

Code	Model	Measuring range	Cable length	Dimensions	Weight
09012016	TK106	-50 to +700°C	27 cm	380 x 100 x 10 mm	0,09 kg


Calibration

Code	Description
09053003	Calibration report ref. Accredia for TK106


TK105 Surface probe

Code	Model	Measuring range	Cable length	Dimensions	Weight
09012020	TK105	-50 to +400°C	27 cm	380 x 100 x 10 mm	0,14 kg


Calibration

Code	Description
09053002	Calibration report ref. Accredia for TK105


TK107 Clamp on probe for measurements on pipes

Code	Model	Measuring range	Cable length	Dimensions	Weight
09012017	TK107	-50 to +120°C	54 cm	230 x 100 x 40 mm	0,16 kg


Calibration

Code	Description
09053004	Calibration report ref. Accredia for TK107



Max opening 36 mm